



**TELEDYNE
INSTRUMENTS**

Advanced Pollution Instrumentation

A Teledyne Technologies Company
9480 Carroll Park Drive, San Diego, CA 92121-5201
Phone (858) 657-9800 Fax: (858) 657-9818 Toll Free 1800 324-5190
E-mail: api-customerservice@teledyne.com <http://www.teledyne-api.com>

Service Note

**98-053 Rev B
2 May, 2007**

UNIT MASS FLOW CONTROLLERS

I. SCOPE:

To inform our customers of a change in mass flow controllers.

II. PARTS:

None.

III. TOOLS:

None.

IV. PROCEDURE:

The purpose of this service note is to inform our customers of a change in mass flow controllers. You might notice that some of our Calibrators have Unit mass flow controllers in them & not the Tylan General mass flow controllers that we have historically used. Be assured that we have done extensive testing on the Unit mass flow controllers & that they are as good as or in some instances better than the Tylan mass flow controllers. The compatibility of the Unit is exactly the same as the Tylan mass flow controller.

If you have questions regarding this procedure or any API equipment, please contact an API Customer Service representative at:

Phone: (619) 675-9800

Fax: (619) 657-9816

Email: customerservice@advpol.com