

CALIBRATION GAS MANUFACTURING:


ENSURING PRECISION IN ENVIRONMENTAL COMPLIANCE

Mark Patten | Vice President Of Technical Gas Services



“The Process is really what you have to do day in and day out to be successful.”

NICK SABAN





|| EXPLORING THE CRUCIAL ROLE OF EPA PROTOCOL CALIBRATION GASES

- EPA Protocol Gases: Essential for Precision-Demand Industries
- Supporting Environmental Compliance Through Gas Manufacturing
- Aiding Emission Reporting in Environmental and Stationary Analysis.
- Participation in EPA's PGVP Mandatory for Manufacturers; 27 Sites Listed as of 2023

OVERVIEW OF CALIBRATION GAS MANUFACTURING

- **Accurately receiving and reentering customer orders**

STEP 1

CONSIDERATIONS:

- Meeting Manufacturer's Blend Tolerances
- Assessing Mixture Component Compatibility
- Correct Cylinder and CGA Valve Identification
- Ensuring Safe Blending
- Product Compliance and Customer Satisfaction

|| RECIPE GENERATION AND CYLINDER PROCESSING

STEP 2

- Create fill recipe: correct cylinder, CGA valve. Recipe generation in grams or kilograms
- Verify cylinder's proper pre-fill processing.
- Evacuation/Bake Out and Passivation steps (Manufacturer Specific)
- Introducing minor components: Direct Add or Base Mixture.
- Daily scale check: NIST-certified, high precision
- Example: Sartorius scales (0.1g floor, 0.01g bench)

CYLINDER SIZE SPECIFICATIONS

Cyl. Size	Nominal Size* Dia X Height (inches)	Nominal* Tare Weight (lbs.)	Water Capacity (lbs.)	Internal Volume @ 70° F (21° C), 1 ATM (liters/cubic feet)		US DOT Specs
300	9.25 X 60	135	110	49.9	1.76	3AA2400
200	9 X 56	115	96	43.8	1.55	3AA2015
80	8.5 X 31	60	37.9	17.2	0.61	3AA2015
40	6 X 24	27	15.2	6.88	0.24	3AA2015
20	4 X 18	12	4.9	2.24	0.08	3AA2015
152	8 X 53	52	64.8	29.5	1.04	3AL2015
82	7.25 X 39	33	34.6	15.7	0.55	3AL2216
32	6.9 X 21	19	13	5.9	0.21	3AL2216
LP239	14.5 X 50	75	238	108	3.83	4BA240
SSB	8 X 37	95	41.6	18.9	0.67	3A1800
10S	4 X 31	21	8.3	3.8	0.13	3A1800
LB	2 X 15	4	1	0.44	0.016	3E1800
XF	12 X 46	180	--	60.9	2.15	8AL
XG	15 X 56	149	278	126.3	4.46	4AA480
XM	10 X 49	90	120	54.3	1.92	3A480
XP	10 X 55	55	124	55.7	1.98	4BA300
QT	3 X 14**	2.5**	2.0	0.900	0.0318	4B-240ET
LP5	12.25 X 18.25	18.5	47.7	21.68	0.76	4BW240

STANDARD CYLINDER SIZES

*Includes 5.5 inches or 4.5 lbs.
for valve and cap

**Includes 4.5 inches or 1.5 lbs.
for valve

EXAMPLE OF RECIPE: 100 PPM NITRIC OXIDE IN NITROGEN

Your Mixture Recipe

Mixture Recipe - 100 PPM NO in Nitrogen

Press. Chart Print \$ Est. Add

Gas	Mass (G)		Moles	Pressure (PSIG)				
	Target	Actual		70 F	80 F	90 F	100 F	110 F
Nitric Oxide	100ppm	0.4929	0.016	-14.5	-14.5	-14.5	-14.5	-14.5
Nitrogen	Balance	4592.9	163.9	2000	2049	2098	2146	2195

EPA Protocol 139.8 CF

Cylinder Style: 152.1S Ten or five year hydrotest: Five

Max. safety device (psi): 3360

Source gas/premix Cylinder: CC209360
Add 45.96 grams source gas/premix to the final mix.

This source gas/premix will contribute the following grams to each component:
Nitric Oxide 0.4929
Nitrogen 45.47

----- SOURCE GAS / PREMIX CALCULATIONS (2 of 2) -----

Use source gas/premix: zNitrogen
Source gas/premix Lot Number: C17L041
Source gas/premix Cylinder: CC191640
Add 4547.3 grams source gas/premix to the final mix.

This source gas/premix will contribute the following grams to each component:
Nitrogen 4547.3

The following additional grams are required after this source gas/premix is added:
Nitric Oxide -0
Nitrogen 0

The recommended valve outlet connection is CGA 660.

Verify the following:
☐ Cylinder Style: 152.1S
☐ DOT Specification: 3AL - DOT Service Pressure: 2015 psig
☐ Dimensions: 8 inches (diameter) by 52 inches (height) - Water Capacity: 1.04 cubic feet

Version 13.95 - 6/19/2020

The mixture fill pressure may not exceed the working pressure of the cylinder.

Mixture Statistics: Specific Volume = 13.80512 cf/lb
Specific Gravity = 0.9673207 relative to air.

[Scroll down for additional information](#)

FILLING PROCESS

STEP 3

- Begin filling process
 - Connect selected cylinders to fill system
 - Evacuate system to remove impurities
 - Add minor and balance components as per recipe
 - Component addition: Direct or base-mixture methods
 - Repeat steps for each component addition
 - Post-fill: Roll or shake for mixture homogeneity
 - Prepare filled cylinders for testing process

|| BAKE OUT SYSTEM



BLOW DOWN PREP MANIFOLD






FILLING SYSTEMS







|| CYLINDER INVERTER

SIX CYLINDER
CAPACITY





MULTI-POINT CALIBRATION FUNDAMENTALS


- Calibration every 30 days for precision
- Adjusts for operational range changes
- Daily 5% relative difference checks
- Monthly recalibration with new curves
- Ensures instrument adapts to range shifts
- Daily checks maintain ongoing accuracy



TESTING PROCESS


- Initiate standardized testing protocols
- Conducting leak detection tests
- Performing gas mixture analysis
- Verifying compliance with quality standards
- Cross-checking with recipe specifications
- Recording test results for traceability
- Ensuring consistency in repeat tests





K 22
0

BANK #1
SHARED





TESTING AND CERTIFICATION

- 4-day incubation period for samples.
- Three assays against NIST Reference Material/equivalent
- Requirement for valid multi-point analysis
- Daily span checks for validity.
- Utilizing TOST statistical model
- 7-day (11 total days) stability period for reactive gases
- Overlap in results leads to cylinder certification
- Non-overlap: Repeat process and certify/disqualify
- Assign expiration date, label, and certification paperwork

ANALYZER BANKS WITH FLOW CONTROL SYSTEMS



SEPARATE LAB SETUPS

MICRO LAB DESIGN









EPA TRACEABILITY PROTOCOL FOR ASSAY AND CERTIFICATION OF GASEOUS CALIBRATION STANDARDS

- Overview of EPA Green Book Protocols
- Assay Procedures for Gaseous Calibration Standards
- Certification Processes under EPA Guidelines
- Ensuring Traceability in Gas Calibration
- Adherence to Environmental Compliance Standards
- Quality Assurance in Calibration Gas Production
- Documentation and Record-Keeping for EPA Certification


MULTIPOINT FLOW DIAGRAM



MULTIPOINT CALIBRATION CURVE



ASSAY FLOW DIAGRAM



ASSAY RESULTS

In this sheet the results of two or three Assays are entered. Calibration dates are entered so Assays having the same calibration uncertainty may be treated correctly. (Assays having a common calibration share the same calibration uncertainty.)

Enter the results for up to three separate assays in chronological order below.

ASSAY 1

29.6437	= estimated concentration
0.52%	= Expanded uncertainty (expressed as percentage of estimated concentration)
0.27	= portion of expanded uncertainty due to calibration
3	= number of measurements
2/13/2023	= date of prior calibration

ASSAY 2

29.9759	= estimated concentration
0.39%	= Expanded uncertainty (expressed as percentage of estimated concentration)
0.28	= portion of expanded uncertainty due to calibration
3	= number of measurements
3/1/2023	= date of prior calibration

ASSAY 3 (if applicable)

29.9629	= estimated concentration
0.46%	= Expanded uncertainty (expressed as percentage of estimated concentration)
1.00	= portion of expanded uncertainty due to calibration
3	= number of measurements
3/1/2023	= date of prior calibration

Number of different calibrations represented by the above data:

N = 2 (If this value seems to be incorrect, check the dates entered for the three assays. The earliest data should be for Assay 1. Trailing spaces may cause N's formula to interpret identical dates as different.)

COMPARISONS

Assay 1 vs Assay 2

				Two One-Sided Equivalence with Assay 1		Variance Components		Total
Calibration	Assay	Lower Confidence	Upper Confidence	Lower Test Significance	Upper Test Significance	Calibration	Imprecision	
1	1					0.0063888	0.017106	0.0234948
2	2	0.2132921	0.4512181	TRUE	FALSE	0.0039172	0.0099554	0.0138726

FALSE indicates an inconsistency where the observed confidence interval of the difference is beyond the tolerance level.

Assay 1 vs Assay 3

				Two One-Sided Equivalence with Assay 1		Variance Components		Total
Calibration	Assay	Lower Confidence	Upper Confidence	Lower Test Significance	Upper Test Significance	Calibration	Imprecision	
1	1					0.0063888	0.017106	0.0234948
2	3	0.1927886	0.445546	TRUE	FALSE	0.0186746	1.868E-06	0.0186765

FALSE indicates an inconsistency where the observed confidence interval of the difference is beyond the tolerance level.

Nothing will appear here if no data have been entered for Assay 3.

Assay 2 vs Assay 3

				Two One-Sided Equivalence with Assay 1		Variance Components		Total
Calibration	Assay	Lower Confidence	Upper Confidence	Lower Test Significance	Upper Test Significance	Calibration	Imprecision	
2	2					0.0039172	0.0099554	0.0138726
2	3	-0.124116	0.0979408	TRUE	TRUE	0.0186746	1.868E-06	0.0186765

FALSE indicates an inconsistency where the observed confidence interval of the difference is beyond the tolerance level.

Nothing will appear here if no data have been entered for Assay 3.

OVERALL ESTIMATE

Case = 15 (right click to see comment)


The standard error of the estimate produced in an assay is equal to approximately 1/2 of the "95% uncertainty." The inverse of the square of the standard error is the (raw) weighting factor used in producing an overall estimate of the concentration. The raw weights are adjusted (Adj. Wt.) so their sum is 1.00.

Calibration	Estimate	Exp. Uncert.	Raw Wt.	Adj. Wt.	Wt.*Conc.	Variance of Wt.*Est.
1	29.643695	0.517%	37401.80627	0.320	9.487	0.00241
2	29.962864	0.355%	79464.42242	0.680	20.374	0.00522
			0	0.000	0.000	0.00000

29.8607174 = overall estimate of the candidate standard's concentration
0.087 = Expanded uncertainty (concentration units)
0.293% = Expanded relative uncertainty

CYLINDER PROGRESSION STEPS

Full Process in View



Q/C & Release

Review & Release.

Fill and Certification

Fill & Analyzation.

Cylinder Processing

Bake Out & Evacuation,
Passivation Treatment and
Pickling.


Cylinder Selection

Select appropriate cylinder and
valve type.

2008 EPA IG Audit



2008 EPA Inspector General Audit




There were 261
components
audited with 28
failures.

Total Failure Rate
of 11%

2010 AUDIT RESULTS



2010 Emission PGVP Results




The results of the 2010 audit are generally consistent with previous audits- a 10% failure rate over all components analyzed, with 40% of the production sites failing at least one gaseous component

2013 LATEST AUDIT RESULTS




2013 Emission PGVP Results



There were 3
producer who
failed and 4
components.


2015 LATEST AUDIT RESULTS



There was another 11% failure rate. Worst Audit Result to date.

2018 LATEST AUDIT RESULTS

2018 PGVP Audit Results



There were 66 cylinders audited with 6 failures.

Total Failure Rate of 9.1%

OBSTACLES IN CALIBRATION GAS MANUFACTURING

- Precise gas mixture creation challenges
- Managing instrument variability and sensor drift
- Adapting to regulatory changes and updates (NO_x certification vs NIST traceability)
- Limited availability of NIST SRMs and equivalents
- Balancing cost and time efficiency in production
- High demand vs. production scalability issues

SRMs AVAILABLE

Gas	Concentration	UC	Available
CO2/N2	0.5 %	Yes	
CO2/N2	1 %	Yes	
CO2/N2	1.5 %	Yes	
CO2/N2	2 %	Yes	
CO2/N2	7 %	Yes	
CO2/N2	16 %	Yes	
CO2/N2	3 %	No	
CO2/N2	3.5 %	No	
CO2/N2	500 PPM	No	
CO/N2	25 PPM	Yes	
CO/N2	50 PPM	Yes	
CO/N2	500 PPM	Yes	
CO/N2	1000 PPM	Yes	
CO/Air	20 PPM	No	
CO/Air	45 PPM	No	
CO/N2	13 PPM	No	
CO/N2	10 PPM	No	
CO/N2	25 PPM	No	
CO/N2	250 PPM	No	
CO/N2	2500 PPM	No	
CO/N2		No	
CO/N2	5000 PPM	No	
CO/N2	1000 PPM	No	
CO/N2	1 %	No	
CO/N2	2 %	No	
CO/N2	4 %	No	
CO/N2	8 %	No	
Methane/Air	1 PPM	No	
Methane/Air	10 PPM	No	
Methane/Air	50 PPM	No	
Methane/Air	100 PPM	No	

NO/N2	1000 PPM	Yes
NO/N2	5 PPM	No
NO/N2	10 PPM	No
NO/N2	20 PPM	No
NO/N2	50 PPM	No
NO/N2	250 PPM	No
NO/N2	500 PPM	No
NO/N2	800 PPM	No
NO/N2	1500 PPM	No
NO/N2	3000 PPM	No
NO2/Air	100 PPM	No
O2/N2	2 %	Yes
O2/N2	10 %	Yes
O2/N2	21 %	Yes
SO2/N2	100 PPM	Yes
SO2/N2	500 PPM	Yes
SO2/N2	1000 PPM	Yes
SO2/N2	1500 PPM	Yes
SO2/N2	3500 PPM	Yes
SO2/N2	5 PPM	No
SO2/N2	50 PPM	No
SO2/N2	2500 PPM	No
Propane/Air	10 PPM	Yes
Propane/Air	50 PPM	Yes
Propane/Air	2500 PPM	Yes
Propane/Air	0.25 PPM	No
Propane/Air	3 PPM	No
Propane/Air	100 PPM	No
Propane/Air	250 PPM	No
Propane/Air	500 PPM	No

EZ BUTTON APP

Red Ball Oxygen

Easy Button

Cart

Account

Scan Asset

Asset 999999999

110 PPM NO BAL N2 EPA Sz152

Product ID	123448
Product Type	Spec Gas : Spec Mix : 152AL
Asset Serial	EB0099999
Department	Department Name Redacted
Department ID	999999
Use-state	Delivered
Delivered	5 months ago
Status	Active

Printable Docs

Certificate of Authenticity (CoA)

Safety Datasheet (SDS)

Orders

Order More

1

Add to Cart

2 Previous Orders using this App

Inventory

Order Items

Your Orders

File Issue

Red Ball Oxygen

Easy Button

Cart

Account

Scan Asset

Asset Serial	EB0099999
Department	Department Name Redacted
Department ID	999999
Use-state	Delivered
Delivered	5 months ago
Status	Active

Printable Docs

Certificate of Authenticity (CoA)

Safety Datasheet (SDS)

Orders

Order More

1

Add to Cart

2 Previous Orders using this App

Make Changes

Pressure (psi.)

Update

Transfer/Return Asset...

Inventory

Order Items

Your Orders

File Issue

Red Ball Oxygen

Easy Button

Cart

Account

Scan Asset

Order Items

Find Available Items...

Show All Items

121026

Sort By Orders

Find Items

Nitrogen, Zero Ambient, Sz152

121026 Popular Product

78 lbs. ea.

1

Add to Cart

Bernard Ns-1218c Nozzle 1/2"

BERNS1218C Popular Product

1

Add to Cart

Bernard Ns-1218b Nozzle 1/2

BERNS1218B Popular Product

1

Add to Cart

Profax 151026 .035 Drive Roll Kit

PRO151026

1

Add to Cart

Miller 151026 Drive Roll

MIL151026

1

Add to Cart

Inventory

Order Items

Your Orders

File Issue

Continued

Red Ball Oxygen

Easy Button

Cart

Account

Scan Asset

Inventory

Dept. Transfers & Returns

Transfer 2 Items...

[Previous Transfers](#)

Accessible Holders & Items

▼ 2 items: Department Name Redacted

2× [126827] 8.25ppm No 8.25ppm CO BAL N2 EPA

► Sz152

▼ 117 items: Department Name Redacted

21× [126323] 1100 PPM CO 13.75% O2 BAL N2, EPA,

► SZ152

20× [126827] 8.25ppm No 8.25ppm CO BAL N2 EPA

► Sz152

▼ 18× [123448] 110 PPM NO BAL N2 EPA Sz152

000000100

000000200

000000300

000000400

000000500

Inventory

Order Items

Your Orders

File Issue

Red Ball Oxygen

Easy Button

Cart

Account

Scan Asset

File an Issue

Details

Select Type of Issue

File this Issue

We will reply to:
me@nameofcompany.example.com

To update your account's email address on file,
go to [Account Settings](#).

Inventory

Order Items

Your Orders

File Issue

|| Last Step, Choose Wisely!

RED Ball
TGS

|| FINAL QUESTION WE SHOULD ASK

Is the process paving the way
toward the success we
envision?

MARK PATTEN

